

# Proficiency testing for in-house and external measuring stations - results and evaluation

## Proficiency testing scheme

## Volatile organic compounds (VOC) with thermal desorption with own sampling

**12 September 2023**

## Summary of laboratory test results

Sample 1

| Unit                                       | 1,2,3-Trimethylbenzene   |          | 4-Methyl-2-pentanone     |          | alpha-Pinene             |          | Benzaldehyde             |          |
|--|--------------------------|----------|--------------------------|----------|--------------------------|----------|--------------------------|----------|
|  | $\mu\text{g}/\text{m}^3$ | Z score  | $\mu\text{g}/\text{m}^3$ | Z score  | $\mu\text{g}/\text{m}^3$ | Z score  | $\mu\text{g}/\text{m}^3$ | Z score  |
| 1  | 50,1                     | 1,80     | 151,0                    | 0,52     | 49,4                     | 0,38     | 45,1                     | 1,66     |
| 23   | 50,0                     | 1,78     | 155,0                    | 0,80     | 41,0                     | -1,38    | 49,0                     | 2,67 E   |
| 24   | 40,0                     | -0,59    | 146,1                    | 0,18     | 39,3                     | -1,75    | 44,4                     | 1,48     |
| 26   | 42,4                     | -0,01    | 140,0                    | -0,25    | 44,8                     | -0,59    | 37,7                     | -0,25    |
| 39   | 52,5                     | 2,36 CE  | 144,5                    | 0,07     | 46,4                     | -0,26    | 37,4                     | -0,33    |
| 44   | 28,1                     | -3,37 FE | 94,8                     | -3,40 E  | 24,0                     | -4,95 FE | 33,7                     | -1,29    |
| 66   | 42,7                     | 0,06     | 159,0                    | 1,08     | 47,1                     | -0,10    | 43,5                     | 1,25     |
| 95   | 39,7                     | -0,65    | 128,9                    | -1,02    | 52,3                     | 0,99     | 30,4                     | -2,13 E  |
| 126  | 32,1                     | -2,43 E  | 131,9                    | -0,81    | 96,5                     | 10,29 BE | 32,3                     | -1,65    |
| 147  |                          |          |                          |          |                          |          |                          |          |
| 152  | 50,5                     | 1,90     | 165,2                    | 1,51     | 52,4                     | 1,01     | 39,0                     | 0,09     |
| 180  | 41,0                     | -0,34    | 115,0                    | -1,99    | 25,5                     | -4,64 FE | 29,5                     | -2,37 E  |
| 206  | 63,7                     | 5,01 FE  | 191,0                    | 3,30 E   | 80,9                     | 7,01 BE  | 54,4                     | 4,07 FE  |
| 209  | 41,0                     | -0,34    | 146,0                    | 0,17     | 48,0                     | 0,09     | 38,0                     | -0,17    |
| 213  | 40,0                     | -0,58    | 127,9                    | -1,09    | 38,7                     | -1,86    | 33,7                     | -1,28    |
| 228  | 42,6                     | 0,05     |                          |          |                          |          |                          |          |
| 230  | 42,0                     | -0,11    | 150,0                    | 0,45     | 59,0                     | 2,40 E   | 44,5                     | 1,51     |
| 233  | 44,5                     | 0,49     | 147,1                    | 0,25     | 51,9                     | 0,90     |                          |          |
| 235  | 32,0                     | -2,46 E  | 82,0                     | -4,29 FE | 24,0                     | -4,96 FE | 15,0                     | -6,12 BE |
| 242  | 45,3                     | 0,67     | 140,4                    | -0,22    | 46,2                     | -0,29    | 49,2                     | 2,73 E   |
| 265  | 43,0                     | 0,13     | 139,0                    | -0,32    | 46,0                     | -0,33    | 40,0                     | 0,35     |
| 284  | 48,5                     | 1,43     | 157,7                    | 0,98     | 49,9                     | 0,49     | 37,8                     | -0,22    |
| 289  | 33,3                     | -2,17 CE | 127,5                    | -1,12    | 40,7                     | -1,46    | 38,3                     | -0,09    |
| 517  | 51,1                     | 2,04 E   | 168,0                    | 1,70     | 52,0                     | 0,93     | 41,1                     | 0,63     |
| -  | -                        | --       | -                        | --       | -                        | --       | -                        | --       |
| Method                                     | ISO 5725-2               |          | ISO 5725-2               |          | ISO 5725-2               |          | ISO 5725-2               |          |
| Assessment                                 | Z <=2,00                 |          | Z <=2,00                 |          | Z <=2,00                 |          | Z <=2,00                 |          |
| No. of laboratories that submitted results | 23                       |          | 22                       |          | 22                       |          | 21                       |          |

|  | 1,2,3-Trimethylbenzene | Z score | 4-Methyl-2-pentanone | Z score | alpha-Pinene | Z score | Benzaldehyde | Z score |
|--|------------------------|---------|----------------------|---------|--------------|---------|--------------|---------|
| Mean   | 42,5                   |         | 143,6                |         | 47,6         |         | 38,7         |         |
| Reprod. s.d.   | 5,3                    |         | 20,9                 |         | 5,7          |         | 6,2          |         |
| Rel. reproducibility s.d.  | 12,50 %                |         | 14,58 %              |         | 12,00 %      |         | 15,94 %      |         |
| Reference value  | 42,3                   |         | 136,8                |         | 43,0         |         | 36,2         |         |
| Target s.d.  | 4,2                    |         | 14,4                 |         | 4,8          |         | 3,9          |         |
| Rel. target s.d.   | 10,00 %                |         | 10,00 %              |         | 10,00 %      |         | 10,00 %      |         |
| Lower limit of tolerance   | 34,0                   |         | 114,8                |         | 38,1         |         | 30,9         |         |
| Upper limit of tolerance   | 50,9                   |         | 172,3                |         | 57,1         |         | 46,4         |         |
| Type B outliers  |                        |         |                      |         | 2            |         | 1            |         |
| Type C outliers  | 2                      |         |                      |         |              |         |              |         |
| Type F outliers  | 2                      |         | 1                    |         | 3            |         | 1            |         |
| No. of laboratories after elimination of outliers type A-D and F (without laboratories that only gave states but no measured values) | 19                     |         | 21                   |         | 17           |         | 19           |         |
| Explanation of outlier types   |                        |         |                      |         |              |         |              |         |
| A: Single outlier  | Grubbs                 |         |                      |         |              |         |              |         |
| B: Differing laboratory mean   | Grubbs                 |         |                      |         |              |         |              |         |
| C: Excessive laboratory s.d.   | Cochran                |         |                      |         |              |         |              |         |
| D: Excluded manually   |                        |         |                      |         |              |         |              |         |
| E: mean outside tolerance limits   |                        |         |                      |         |              |         |              |         |
| F:  Z-Score >3,50  |                        |         |                      |         |              |         |              |         |

|      | Benzene | Z score  | Cumene | Z score  | n-Decane | Z score  | n-Heptane | Z score  | p-Xylene | Z score  |
|------|---------|----------|--------|----------|----------|----------|-----------|----------|----------|----------|
| Unit | µg/m³   |          | µg/m³  |          | µg/m³    |          | µg/m³     |          | µg/m³    |          |
| 1    | 39,5    | -0,27    | 69,0   | 0,72     | 136,0    | 0,56     | 39,6      | 0,06     | 36,1     | 1,04     |
| 23   | 36,0    | -1,13    | 74,0   | 1,49     | 150,0    | 1,65     | 41,0      | 0,41     | 39,0     | 1,92     |
| 24   | 41,2    | 0,16     | 61,1   | -0,52    | 133,0    | 0,33     | 48,4      | 2,30 E   | 31,5     | -0,38    |
| 26   | 37,4    | -0,79    | 62,7   | -0,26    | 124,5    | -0,33    | 35,8      | -0,92    | 32,5     | -0,05    |
| 39   | 41,5    | 0,23     | 62,1   | -0,35    | 111,5    | -1,34    | 40,9      | 0,38     | 34,3     | 0,47     |
| 44   | 24,7    | -3,91 FE | 35,3   | -4,51 BE | 75,1     | -4,17 FE | 14,6      | -6,30 FE | 12,0     | -6,32 BE |

|  | Benzene    | Z score  | Cumene     | Z score | n-Decane   | Z score | n-Heptane  | Z score  | p-Xylene   | Z score  |
|--|------------|----------|------------|---------|------------|---------|------------|----------|------------|----------|
| 66   | 36,6       | -0,99    | 76,3       | 1,85    | 134,6      | 0,45    | 37,2       | -0,55    | 31,9       | -0,25    |
| 95   | 49,5       | 2,19 E   | 56,3       | -1,25   | 124,8      | -0,31   | 44,2       | 1,23     | 27,7       | -1,52    |
| 126  | 28,3       | -3,03 E  | 67,0       | 0,40    | 113,6      | -1,18   | 30,8       | -2,19 E  | 25,4       | -2,23 E  |
| 147  | 40,0       | -0,15    |            |         |            |         | 45,5       | 1,56     | 35,5       | 0,85     |
| 152  | 43,2       | 0,64     | 65,6       | 0,19    | 136,6      | 0,61    | 38,5       | -0,22    | 36,6       | 1,19     |
| 180  | 22,0       | -4,58 FE | 62,0       | -0,37   | 115,0      | -1,07   | 29,5       | -2,51 E  | 25,5       | -2,20 E  |
| 206  | 58,9       | 4,49 FE  | 92,8       | 4,41 BE | 170,9      | 3,28 E  | 64,4       | 6,36 FE  | 28,4       | -1,33    |
| 209  | 37,0       | -0,89    | 63,0       | -0,22   | 122,0      | -0,53   | 38,0       | -0,35    | 34,0       | 0,40     |
| 213  | 35,3       | -1,32    | 60,5       | -0,60   | 137,7      | 0,70    | 39,1       | -0,06    | 31,2       | -0,46    |
| 228  | 51,9       | 2,78 E   | 72,0       | 1,18    | 144,8      | 1,25    | 63,3       | 6,08 FE  | 37,3       | 1,39     |
| 230  | 47,5       | 1,70     | 60,5       | -0,60   | 122,5      | -0,49   | 52,5       | 3,34 E   | 38,5       | 1,77     |
| 233  | 41,4       | 0,20     | 65,7       | 0,21    | 132,5      | 0,29    | 41,4       | 0,53     | 34,7       | 0,61     |
| 235  | 13,0       | -6,80 FE | 61,0       | -0,53   | 87,0       | -3,24 E | 21,0       | -4,67 FE | 20,0       | -3,88 FE |
| 242  | 39,5       | -0,28    | 65,8       | 0,22    | 131,9      | 0,24    | 37,3       | -0,54    | 31,8       | -0,28    |
| 265  | 38,0       | -0,64    | 62,0       | -0,37   | 138,0      | 0,72    | 38,0       | -0,35    | 36,0       | 1,01     |
| 284  | 41,5       | 0,22     | 63,1       | -0,20   | 128,0      | -0,06   | 36,7       | -0,68    | 35,0       | 0,70     |
| 289  | 36,0       | -1,15    | 63,9       | -0,08   | 113,0      | -1,22   | 40,0       | 0,16 C   | 29,4       | -1,00 C  |
| 517  | 44,2       | 0,89     | 65,9       | 0,23    | 140,0      | 0,87    | 38,5       | -0,22    | 37,0       | 1,31     |
| -  | -          | --       | -          | --      | -          | --      | -          | --       | -          | --       |
| Method                                     | ISO 5725-2 |          | ISO 5725-2 |         | ISO 5725-2 |         | ISO 5725-2 |          | ISO 5725-2 |          |
| Assessment                                 | Z <=2,00   |          | Z <=2,00   |         | Z <=2,00   |         | Z <=2,00   |          | Z <=2,00   |          |
| No. of laboratories that submitted results | 24         |          | 23         |         | 23         |         | 24         |          | 24         |          |
| Mean                                       | 40,6       |          | 64,4       |         | 128,8      |         | 39,4       |          | 32,7       |          |
| Reprod. s.d.                               | 6,2        |          | 4,9        |         | 17,3       |         | 6,2        |          | 4,3        |          |
| Rel. reproducibility s.d.                  | 15,22 %    |          | 7,64 %     |         | 13,47 %    |         | 15,74 %    |          | 13,28 %    |          |
| Reference value                            | 34,5       |          | 60,0       |         | 128,8      |         | 33,9       |          | 29,8       |          |
| Target s.d.                                | 4,1        |          | 6,4        |         | 12,9       |         | 3,9        |          | 3,3        |          |
| Rel. target s.d.                           | 10,00 %    |          | 10,00 %    |         | 10,00 %    |         | 10,00 %    |          | 10,00 %    |          |
| Lower limit of tolerance                   | 32,5       |          | 51,5       |         | 103,0      |         | 31,5       |          | 26,2       |          |
| Upper limit of tolerance                   | 48,7       |          | 77,3       |         | 154,5      |         | 47,2       |          | 39,2       |          |
| Type B outliers                            |            |          | 2          |         |            |         |            |          | 1          |          |
| Type C outliers                            |            |          |            |         |            |         | 1          |          | 1          |          |
| Type F outliers                            | 4          |          |            |         | 1          |         | 4          |          | 1          |          |
| No. of laboratories after elimination of   | 20         |          | 21         |         | 22         |         | 19         |          | 21         |          |

|  | Benzene    | Z score | Cumene  | Z score | n-Decane | Z score | n-Heptane | Z score | p-Xylene | Z score |
|--|------------|---------|---------|---------|----------|---------|-----------|---------|----------|---------|
| outliers type A-D and F (w ithout laboratories that only gave states but no measured values) |            |         |         |         |          |         |           |         |          |         |
|  | Toluene    |         | Z score |         |          |         |           |         |          |         |
| Unit   | µg/m³      |         |         |         |          |         |           |         |          |         |
| 1  | 31,6       | 1,04    |         |         |          |         |           |         |          |         |
| 23   | 32,0       | 1,18    |         |         |          |         |           |         |          |         |
| 24   | 27,4       | -0,41   |         |         |          |         |           |         |          |         |
| 26   | 27,0       | -0,57   |         |         |          |         |           |         |          |         |
| 39   | 30,5       | 0,67    |         |         |          |         |           |         |          |         |
| 44   | 11,0       | -6,15   | BE      |         |          |         |           |         |          |         |
| 66   | 27,7       | -0,33   |         |         |          |         |           |         |          |         |
| 95   | 23,5       | -1,78   |         |         |          |         |           |         |          |         |
| 126  | 19,7       | -3,11   | E       |         |          |         |           |         |          |         |
| 147  | 31,0       | 0,83    |         |         |          |         |           |         |          |         |
| 152  | 31,1       | 0,86    |         |         |          |         |           |         |          |         |
| 180  | 24,0       | -1,62   |         |         |          |         |           |         |          |         |
| 206  | 36,0       | 2,56    | E       |         |          |         |           |         |          |         |
| 209  | 26,0       | -0,92   |         |         |          |         |           |         |          |         |
| 213  | 27,0       | -0,57   |         |         |          |         |           |         |          |         |
| 228  | 33,0       | 1,54    |         |         |          |         |           |         |          |         |
| 230  | 36,5       | 2,75    | E       |         |          |         |           |         |          |         |
| 233  | 29,5       | 0,31    |         |         |          |         |           |         |          |         |
| 235  | 21,0       | -2,67   | E       |         |          |         |           |         |          |         |
| 242  | 26,1       | -0,90   |         |         |          |         |           |         |          |         |
| 265  | 30,0       | 0,48    |         |         |          |         |           |         |          |         |
| 284  | 29,5       | 0,30    |         |         |          |         |           |         |          |         |
| 289  | 28,5       | -0,05   |         |         |          |         |           |         |          |         |
| 517  | 31,2       | 0,90    |         |         |          |         |           |         |          |         |
| -  | -          | --      |         |         |          |         |           |         |          |         |
| Method   | ISO 5725-2 |         |         |         |          |         |           |         |          |         |

---

|  | Toluene  | Z score |
|--|----------|---------|
| Assessment   | Z <=2,00 |         |
| No. of laboratories that submitted results   | 24       |         |
| Mean   | 28,6     |         |
| Reprod. s.d.   | 4,6      |         |
| Rel. reproducibility s.d.  | 16,20 %  |         |
| Reference value  | 24,5     |         |
| Target s.d.  | 2,9      |         |
| Rel. target s.d.   | 10,00 %  |         |
| Lower limit of tolerance   | 22,9     |         |
| Upper limit of tolerance   | 34,4     |         |
| Type B outliers  | 1        |         |
| No. of laboratories after elimination of outliers type A-D and F (without laboratories that only gave states but no measured values) | 23       |         |

---

## Summary of laboratory test results

Sample 2

| Unit                                       | 1,2,3-Trimethylbenzene   |          | 4-Methyl-2-pentanone     |          | alpha-Pinene             |          | Benzaldehyde             |          |
|--|--------------------------|----------|--------------------------|----------|--------------------------|----------|--------------------------|----------|
|  | $\mu\text{g}/\text{m}^3$ | Z score  | $\mu\text{g}/\text{m}^3$ | Z score  | $\mu\text{g}/\text{m}^3$ | Z score  | $\mu\text{g}/\text{m}^3$ | Z score  |
| 1  | 92,0                     | 1,07     | 43,7                     | 1,15     | 89,8                     | 0,40     | 31,8                     | 1,88     |
| 23   | 99,0                     | 1,91     | 41,0                     | 0,46     | 97,0                     | 1,23     | 34,0                     | 2,70 E   |
| 24   | 77,0                     | -0,73    | 37,9                     | -0,32    | 78,5                     | -0,92    | 28,2                     | 0,53     |
| 26   | 78,4                     | -0,57    | 39,2                     | 0,01     | 83,8                     | -0,29    | 26,0                     | -0,29    |
| 39   | 89,0                     | 0,71     | 42,6                     | 0,87     | 61,5                     | -2,88 E  | 30,6                     | 1,43     |
| 44   | 85,7                     | 0,32     | 36,7                     | -0,62    | 89,4                     | 0,35     | 34,7                     | 2,96 E   |
| 66   | 78,8                     | -0,52    | 42,3                     | 0,80     | 86,5                     | 0,01     | 28,9                     | 0,79     |
| 95   | 80,4                     | -0,33    | 34,3                     | -1,25    | 97,2                     | 1,26     | 21,2                     | -2,10 E  |
| 126  | 74,2                     | -1,07 C  | 34,5                     | -1,20    | 153,2                    | 7,73 FE  | 21,3                     | -2,06 E  |
| 147  |                          |          |                          |          |                          |          |                          |          |
| 152  | 87,9                     | 0,58     | 41,3                     | 0,54     | 89,2                     | 0,33     | 26,4                     | -0,14    |
| 180  | 80,5                     | -0,31    | 35,5                     | -0,94    | 44,0                     | -4,91 FE | 21,5                     | -1,97    |
| 206  | 122,7                    | 4,77 BE  | 47,9                     | 2,21 E   | 156,6                    | 8,13 FE  | 39,8                     | 4,84 FE  |
| 209  | 75,0                     | -0,97    | 42,0                     | 0,72     | 86,0                     | -0,04    | 25,0                     | -0,66    |
| 213  | 75,8                     | -0,88    | 35,5                     | -0,93    | 74,3                     | -1,40    | 22,3                     | -1,69    |
| 228  | 92,7                     | 1,16     |                          |          |                          |          |                          |          |
| 230  | 75,5                     | -0,91    | 41,5                     | 0,59     | 106,5                    | 2,33 E   | 28,5                     | 0,64     |
| 233  | 80,6                     | -0,30    | 41,5                     | 0,59     | 88,1                     | 0,20     |                          |          |
| 235  | 54,5                     | -3,44 BE | 22,5                     | -4,26 BE | 43,0                     | -5,02 FE | 7,3                      | -7,26 BE |
| 242  | 83,4                     | 0,04     | 38,6                     | -0,13    | 85,8                     | -0,07    | 33,0                     | 2,30 E   |
| 265  | 78,0                     | -0,61    | 39,0                     | -0,05    | 87,0                     | 0,07     | 27,0                     | 0,08     |
| 284  | 87,9                     | 0,58     | 41,2                     | 0,52     | 89,7                     | 0,38     | 26,9                     | 0,05     |
| 289  | 60,2                     | -2,76 CE | 33,1                     | -1,54    | 66,6                     | -2,29 E  | 24,9                     | -0,70    |
| 517  | 88,8                     | 0,69     | 42,2                     | 0,77     | 89,8                     | 0,40     | 27,9                     | 0,42     |
| -  | -                        | --       | -                        | --       | -                        | --       | -                        | --       |
| Method                                     | ISO 5725-2               |          | ISO 5725-2               |          | ISO 5725-2               |          | ISO 5725-2               |          |
| Assessment                                 | Z <=2,00                 |          | Z <=2,00                 |          | Z <=2,00                 |          | Z <=2,00                 |          |
| No. of laboratories that submitted results | 23                       |          | 22                       |          | 22                       |          | 21                       |          |

|  | 1,2,3-Trimethylbenzene | Z score | 4-Methyl-2-pentanone | Z score | alpha-Pinene | Z score | Benzaldehyde | Z score |
|--|------------------------|---------|----------------------|---------|--------------|---------|--------------|---------|
| Mean   | 83,1                   |         | 39,2                 |         | 86,4         |         | 26,8         |         |
| Reprod. s.d.   | 6,7                    |         | 4,2                  |         | 11,5         |         | 4,4          |         |
| Rel. reproducibility s.d.  | 8,12 %                 |         | 10,62 %              |         | 13,30 %      |         | 16,34 %      |         |
| Reference value  | 86,1                   |         | 36,8                 |         | 80,1         |         | 24,3         |         |
| Target s.d.  | 8,3                    |         | 3,9                  |         | 8,6          |         | 2,7          |         |
| Rel. target s.d.   | 10,00 %                |         | 10,00 %              |         | 10,00 %      |         | 10,00 %      |         |
| Lower limit of tolerance   | 66,5                   |         | 31,3                 |         | 69,1         |         | 21,4         |         |
| Upper limit of tolerance   | 99,7                   |         | 47,0                 |         | 103,7        |         | 32,1         |         |
| Type B outliers  | 2                      |         | 1                    |         |              |         | 1            |         |
| Type C outliers  | 2                      |         |                      |         |              |         |              |         |
| Type F outliers  |                        |         |                      |         | 4            |         | 1            |         |
| No. of laboratories after elimination of outliers type A-D and F (without laboratories that only gave states but no measured values) | 19                     |         | 21                   |         | 18           |         | 19           |         |
| Explanation of outlier types   |                        |         |                      |         |              |         |              |         |
| A: Single outlier  | Grubbs                 |         |                      |         |              |         |              |         |
| B: Differing laboratory mean   | Grubbs                 |         |                      |         |              |         |              |         |
| C: Excessive laboratory s.d.   | Cochran                |         |                      |         |              |         |              |         |
| D: Excluded manually   |                        |         |                      |         |              |         |              |         |
| E: mean outside tolerance limits   |                        |         |                      |         |              |         |              |         |
| F:  Z-Score >3,50  |                        |         |                      |         |              |         |              |         |

|      | Benzene | Z score | Cumene | Z score | n-Decane | Z score | n-Heptane | Z score | p-Xylene | Z score |
|------|---------|---------|--------|---------|----------|---------|-----------|---------|----------|---------|
| Unit | µg/m³   |         | µg/m³  |         | µg/m³    |         | µg/m³     |         | µg/m³    |         |
| 1    | 86,9    | 0,46    | 53,7   | 0,43    | 85,9     | 1,09    | 77,2      | 0,55    | 64,0     | 0,85    |
| 23   | 101,0   | 2,16 E  | 63,0   | 2,23 E  | 97,0     | 2,52 E  | 92,0      | 2,57 E  | 74,0     | 2,55 E  |
| 24   | 92,7    | 1,17    | 47,1   | -0,86   | 79,7     | 0,28    | 95,1      | 2,99 E  | 57,9     | -0,19   |
| 26   | 84,3    | 0,15    | 49,5   | -0,39   | 77,3     | -0,02   | 70,5      | -0,36   | 59,9     | 0,16    |
| 39   | 65,3    | -2,14 E | 51,2   | -0,06   | 71,0     | -0,84   | 71,8      | -0,19   | 61,6     | 0,44    |
| 44   | 67,6    | -1,86   | 63,1   | 2,26 E  | 63,0     | -1,87   | 62,7      | -1,43   | 50,7     | -1,40   |



|  | Benzene    | Z score  | Cumene     | Z score | n-Decane   | Z score | n-Heptane  | Z score  | p-Xylene   | Z score  |
|--|------------|----------|------------|---------|------------|---------|------------|----------|------------|----------|
| 66   | 90,6       | 0,91     | 58,5       | 1,36    | 76,4       | -0,14   | 73,0       | -0,02    | 59,2       | 0,04     |
| 95   | 90,3       | 0,88     | 43,1       | -1,64   | 73,9       | -0,46   | 82,1       | 1,22     | 51,9       | -1,21    |
| 126  | 64,1       | -2,28 E  | 58,8       | 1,42    | 70,2       | -0,94   | 59,9       | -1,82    | 50,8       | -1,38    |
| 147  | 63,0       | -2,41 E  |            |         |            |         | 65,0       | -1,12    | 65,0       | 1,02     |
| 152  | 88,7       | 0,68     | 52,7       | 0,23    | 78,2       | 0,09    | 73,8       | 0,09     | 62,7       | 0,63     |
| 180  | 49,5       | -4,04 FE | 52,5       | 0,19    | 76,0       | -0,19   | 60,5       | -1,73    | 49,0       | -1,69    |
| 206  | 131,1      | 5,78 FE  | 71,9       | 3,96 BE | 100,5      | 2,98 E  | 130,4      | 7,83 FE  | 52,8       | -1,06    |
| 209  | 79,0       | -0,49    | 49,0       | -0,49   | 75,0       | -0,32   | 74,0       | 0,11     | 61,0       | 0,34     |
| 213  | 78,3       | -0,57    | 48,1       | -0,67   | 83,8       | 0,81    | 79,3       | 0,84     | 58,3       | -0,12    |
| 228  | 97,3       | 1,72     | 53,8       | 0,45    | 84,9       | 0,96    | 109,3      | 4,93 FE  | 61,6       | 0,45     |
| 230  | 117,5      | 4,15 FE  | 48,5       | -0,58   | 76,5       | -0,13   | 118,5      | 6,20 FE  | 72,5       | 2,29 E   |
| 233  | 89,7       | 0,81     | 52,2       | 0,13    | 85,0       | 0,96    | 80,0       | 0,94     | 62,1       | 0,53     |
| 235  | 31,0       | -6,27 FE | 47,5       | -0,78   | 52,0       | -3,29 E | 45,5       | -3,78 FE | 37,0       | -3,73 FE |
| 242  | 83,2       | 0,01     | 50,9       | -0,12   | 78,5       | 0,13    | 70,6       | -0,35    | 58,5       | -0,08    |
| 265  | 84,0       | 0,12     | 48,0       | -0,68   | 85,0       | 0,97    | 76,0       | 0,39     | 64,0       | 0,85     |
| 284  | 89,9       | 0,83     | 52,6       | 0,21    | 78,2       | 0,09    | 74,0       | 0,11     | 62,6       | 0,61     |
| 289  | 76,3       | -0,82    | 47,7       | -0,74   | 69,4       | -1,04   | 75,9       | 0,37 C   | 53,5       | -0,93    |
| 517  | 91,5       | 1,02     | 52,9       | 0,27    | 80,2       | 0,35    | 74,9       | 0,24     | 63,3       | 0,73     |
| -  | -          | --       | -          | --      | -          | --      | -          | --       | -          | --       |
| Method                                     | ISO 5725-2 |          | ISO 5725-2 |         | ISO 5725-2 |         | ISO 5725-2 |          | ISO 5725-2 |          |
| Assessment                                 | Z <=2,00   |          | Z <=2,00   |         | Z <=2,00   |         | Z <=2,00   |          | Z <=2,00   |          |
| No. of laboratories that submitted results | 24         |          | 23         |         | 23         |         | 24         |          | 24         |          |
| Mean                                       | 83,0       |          | 51,5       |         | 77,5       |         | 73,2       |          | 59,0       |          |
| Reprod. s.d.                               | 11,4       |          | 5,1        |         | 10,9       |         | 9,6        |          | 6,9        |          |
| Rel. reproducibility s.d.                  | 13,78 %    |          | 9,97 %     |         | 14,02 %    |         | 13,09 %    |          | 11,70 %    |          |
| Reference value                            | 82,7       |          | 47,6       |         | 79,4       |         | 70,6       |          | 59,2       |          |
| Target s.d.                                | 8,3        |          | 5,2        |         | 7,7        |         | 7,3        |          | 5,9        |          |
| Rel. target s.d.                           | 10,00 %    |          | 10,00 %    |         | 10,00 %    |         | 10,00 %    |          | 10,00 %    |          |
| Lower limit of tolerance                   | 66,4       |          | 41,2       |         | 62,0       |         | 58,5       |          | 47,2       |          |
| Upper limit of tolerance                   | 99,6       |          | 61,8       |         | 93,0       |         | 87,8       |          | 70,8       |          |
| Type B outliers                            |            |          | 1          |         |            |         |            |          |            |          |
| Type C outliers                            |            |          |            |         |            |         | 1          |          |            |          |
| Type F outliers                            | 4          |          |            |         |            |         | 4          |          | 1          |          |
| No. of laboratories after elimination of   | 20         |          | 22         |         | 23         |         | 19         |          | 23         |          |

|  | Benzene    | Z score | Cumene  | Z score | n-Decane | Z score | n-Heptane | Z score | p-Xylene | Z score |
|--|------------|---------|---------|---------|----------|---------|-----------|---------|----------|---------|
| outliers type A-D and F (w ithout laboratories that only gave states but no measured values) |            |         |         |         |          |         |           |         |          |         |
|  | Toluene    |         | Z score |         |          |         |           |         |          |         |
| Unit   | µg/m³      |         |         |         |          |         |           |         |          |         |
| 1  | 52,3       | 0,71    |         |         |          |         |           |         |          |         |
| 23   | 56,0       | 1,46    |         |         |          |         |           |         |          |         |
| 24   | 48,6       | -0,04   |         |         |          |         |           |         |          |         |
| 26   | 48,6       | -0,04   |         |         |          |         |           |         |          |         |
| 39   | 52,4       | 0,73    |         |         |          |         |           |         |          |         |
| 44   | 41,2       | -1,57   |         |         |          |         |           |         |          |         |
| 66   | 50,4       | 0,32    |         |         |          |         |           |         |          |         |
| 95   | 42,1       | -1,39   |         |         |          |         |           |         |          |         |
| 126  | 36,9       | -2,44 E |         |         |          |         |           |         |          |         |
| 147  | 38,0       | -2,22 E |         |         |          |         |           |         |          |         |
| 152  | 53,0       | 0,85    |         |         |          |         |           |         |          |         |
| 180  | 46,0       | -0,58   |         |         |          |         |           |         |          |         |
| 206  | 60,8       | 2,45 E  |         |         |          |         |           |         |          |         |
| 209  | 46,0       | -0,58   |         |         |          |         |           |         |          |         |
| 213  | 49,8       | 0,20    |         |         |          |         |           |         |          |         |
| 228  | 51,8       | 0,60    |         |         |          |         |           |         |          |         |
| 230  | 65,0       | 3,31 E  |         |         |          |         |           |         |          |         |
| 233  | 52,1       | 0,66    |         |         |          |         |           |         |          |         |
| 235  | 36,0       | -2,63 E |         |         |          |         |           |         |          |         |
| 242  | 45,0       | -0,78   |         |         |          |         |           |         |          |         |
| 265  | 53,0       | 0,85    |         |         |          |         |           |         |          |         |
| 284  | 52,7       | 0,79    |         |         |          |         |           |         |          |         |
| 289  | 48,5       | -0,06   |         |         |          |         |           |         |          |         |
| 517  | 53,6       | 0,97    |         |         |          |         |           |         |          |         |
| -  | -          | --      |         |         |          |         |           |         |          |         |
| Method   | ISO 5725-2 |         |         |         |          |         |           |         |          |         |


---

|  | Toluene | Z score         |
|--|---------|-----------------|
| Assessment   |         | $ Z  \leq 2,00$ |
| No. of laboratories that submitted results   |         | 24              |
| Mean   |         | 48,8            |
| Reprod. s.d.   |         | 7,8             |
| Rel. reproducibility s.d.  |         | 15,97 %         |
| Reference value  |         | 47,6            |
| Target s.d.  |         | 4,9             |
| Rel. target s.d.   |         | 10,00 %         |
| Lower limit of tolerance   |         | 39,1            |
| Upper limit of tolerance   |         | 58,6            |
| No. of laboratories after elimination of outliers type A-D and F (without laboratories that only gave states but no measured values) |         | 24              |

---


## Summary results

|                  |                        |                     |   |
|------------------|------------------------|---------------------|---|
| Measurand:       | 1,2,3-Trimethylbenzene | Mean:               | 42,5 µg/m <sup>3</sup>                            |
| Sample:          | 1                      | Reprod. s.d.:       | 5,3 µg/m <sup>3</sup>                             |
| Method:          | ISO 5725-2             | Rel.reprod. s.d.:   | 12,50%  |
| Rel.target s.d.: | 10,00%                 | Reference value:    | 42,3 µg/m <sup>3</sup>                            |
|                  |                        | Range of tolerance: | 34,0 - 50,9 µg/m <sup>3</sup> ( Z-Score  <= 2,00) |




## Summary results

|                  |                      |                     |   |
|------------------|----------------------|---------------------|---|
| Measurand:       | 4-Methyl-2-pentanone | Mean:               | 143,6 µg/m <sup>3</sup>                             |
| Sample:          | 1                    | Reprod. s.d.:       | 20,9 µg/m <sup>3</sup>                              |
| Method:          | ISO 5725-2           | Rel.reprod. s.d.:   | 14,58%  |
| Rel.target s.d.: | 10,00%               | Reference value:    | 136,8 µg/m <sup>3</sup>                             |
| :                |                      | Range of tolerance: | 114,8 - 172,3 µg/m <sup>3</sup> ( Z-Score  <= 2,00) |




## Summary results

|                         |                     |                            |   |
|-------------------------|---------------------|----------------------------|---|
| <b>Measurand:</b>       | <b>alpha-Pinene</b> | <b>Mean:</b>               | <b>47,6 µg/m³</b>                           |
| <b>Sample:</b>          | <b>1</b>            | <b>Reprod. s.d.:</b>       | <b>5,7 µg/m³</b>                            |
| <b>Method:</b>          | <b>ISO 5725-2</b>   | <b>Rel.reprod. s.d.:</b>   | <b>12,00%</b>                               |
| <b>Rel.target s.d.:</b> | <b>10,00%</b>       | <b>Reference value:</b>    | <b>43,0 µg/m³</b>                           |
|                         |                     | <b>Range of tolerance:</b> | <b>38,1 - 57,1 µg/m³ ( Z-Score  ≤ 2,00)</b> |




## Summary results

|                  |              |                     |                                       |
|------------------|--------------|---------------------|---------------------------------------|
| Measurand:       | Benzaldehyde | Mean:               | 38,7 µg/m³                            |
| Sample:          | 1            | Reprod. s.d.:       | 6,2 µg/m³                             |
| Method:          | ISO 5725-2   | Rel.reprod. s.d.:   | 15,94%                                |
| Rel.target s.d.: | 10,00%       | Reference value:    | 36,2 µg/m³                            |
|                  |              | Range of tolerance: | 30,9 - 46,4 µg/m³ ( Z-Score  <= 2,00) |



## Summary results


|                         |                   |                            |   |
|-------------------------|-------------------|----------------------------|---|
| <b>Measurand:</b>       | <b>Benzene</b>    | <b>Mean:</b>               | <b>40,6 µg/m³</b>                           |
| <b>Sample:</b>          | <b>1</b>          | <b>Reprod. s.d.:</b>       | <b>6,2 µg/m³</b>                            |
| <b>Method:</b>          | <b>ISO 5725-2</b> | <b>Rel.reprod. s.d.:</b>   | <b>15,22%</b>                               |
| <b>Rel.target s.d.:</b> | <b>10,00%</b>     | <b>Reference value:</b>    | <b>34,5 µg/m³</b>                           |
|                         |                   | <b>Range of tolerance:</b> | <b>32,5 - 48,7 µg/m³ ( Z-Score  ≤ 2,00)</b> |






## Summary results

|                  |            |                     |   |
|------------------|------------|---------------------|---|
| Measurand:       | Cumene     | Mean:               | 64,4 µg/m <sup>3</sup>                            |
| Sample:          | 1          | Reprod. s.d.:       | 4,9 µg/m <sup>3</sup>                             |
| Method:          | ISO 5725-2 | Rel.reprod. s.d.:   | 7,64%   |
| Rel.target s.d.: | 10,00%     | Reference value:    | 60,0 µg/m <sup>3</sup>                            |
|                  |            | Range of tolerance: | 51,5 - 77,3 µg/m <sup>3</sup> ( Z-Score  <= 2,00) |




## Summary results

|                  |            |                     |   |
|------------------|------------|---------------------|---|
| Measurand:       | n-Decane   | Mean:               | 128,8 µg/m <sup>3</sup>                             |
| Sample:          | 1          | Reprod. s.d.:       | 17,3 µg/m <sup>3</sup>                              |
| Method:          | ISO 5725-2 | Rel.reprod. s.d.:   | 13,47%  |
| Rel.target s.d.: | 10,00%     | Reference value:    | 128,8 µg/m <sup>3</sup>                             |
| :                |            | Range of tolerance: | 103,0 - 154,5 µg/m <sup>3</sup> ( Z-Score  <= 2,00) |




## Summary results

|                         |                   |                            |   |
|-------------------------|-------------------|----------------------------|---|
| <b>Measurand:</b>       | <b>n-Heptane</b>  | <b>Mean:</b>               | <b>39,4 µg/m³</b>                               |
| <b>Sample:</b>          | <b>1</b>          | <b>Reprod. s.d.:</b>       | <b>6,2 µg/m³</b>                                |
| <b>Method:</b>          | <b>ISO 5725-2</b> | <b>Rel.reprod. s.d.:</b>   | <b>15,74%</b>                                   |
| <b>Rel.target s.d.:</b> | <b>10,00%</b>     | <b>Reference value:</b>    | <b>33,9 µg/m³</b>                               |
|                         |                   | <b>Range of tolerance:</b> | <b>31,5 - 47,2 µg/m³ ( Z-Score  &lt;= 2,00)</b> |




## Summary results

|                  |            |                     |   |
|------------------|------------|---------------------|---|
| Measurand:       | p-Xylene   | Mean:               | 32,7 µg/m <sup>3</sup>                            |
| Sample:          | 1          | Reprod. s.d.:       | 4,3 µg/m <sup>3</sup>                             |
| Method:          | ISO 5725-2 | Rel.reprod. s.d.:   | 13,28%  |
| Rel.target s.d.: | 10,00%     | Reference value:    | 29,8 µg/m <sup>3</sup>                            |
|                  |            | Range of tolerance: | 26,2 - 39,2 µg/m <sup>3</sup> ( Z-Score  <= 2,00) |




## Summary results

|                         |                   |                            |   |
|-------------------------|-------------------|----------------------------|---|
| <b>Measurand:</b>       | <b>Toluene</b>    | <b>Mean:</b>               | <b>28,6 µg/m³</b>                               |
| <b>Sample:</b>          | <b>1</b>          | <b>Reprod. s.d.:</b>       | <b>4,6 µg/m³</b>                                |
| <b>Method:</b>          | <b>ISO 5725-2</b> | <b>Rel.reprod. s.d.:</b>   | <b>16,20%</b>                                   |
| <b>Rel.target s.d.:</b> | <b>10,00%</b>     | <b>Reference value:</b>    | <b>24,5 µg/m³</b>                               |
|                         |                   | <b>Range of tolerance:</b> | <b>22,9 - 34,4 µg/m³ ( Z-Score  &lt;= 2,00)</b> |




# Sample chart of Z-scores

Sample 1




## Summary results

|                  |                        |                     |                                      |
|------------------|------------------------|---------------------|--------------------------------------|
| Measurand:       | 1,2,3-Trimethylbenzene | Mean:               | 83,1 µg/m³                           |
| Sample:          | 2                      | Reprod. s.d.:       | 6,7 µg/m³                            |
| Method:          | ISO 5725-2             | Rel.reprod. s.d.:   | 8,12%                                |
| Rel.target s.d.: | 10,00%                 | Reference value:    | 86,1 µg/m³                           |
|                  |                        | Range of tolerance: | 66,5 - 99,7 µg/m³ ( Z-Score  ≤ 2,00) |



## Summary results


|                         |                             |                            |   |
|-------------------------|-----------------------------|----------------------------|---|
| <b>Measurand:</b>       | <b>4-Methyl-2-pentanone</b> | <b>Mean:</b>               | <b>39,2 µg/m³</b>                               |
| <b>Sample:</b>          | <b>2</b>                    | <b>Reprod. s.d.:</b>       | <b>4,2 µg/m³</b>                                |
| <b>Method:</b>          | <b>ISO 5725-2</b>           | <b>Rel.reprod. s.d.:</b>   | <b>10,62%</b>                                   |
| <b>Rel.target s.d.:</b> | <b>10,00%</b>               | <b>Reference value:</b>    | <b>36,8 µg/m³</b>                               |
|                         |                             | <b>Range of tolerance:</b> | <b>31,3 - 47,0 µg/m³ ( Z-Score  &lt;= 2,00)</b> |






## Summary results

|                  |              |                     |                                       |
|------------------|--------------|---------------------|---------------------------------------|
| Measurand:       | alpha-Pinene | Mean:               | 86,4 µg/m³                            |
| Sample:          | 2            | Reprod. s.d.:       | 11,5 µg/m³                            |
| Method:          | ISO 5725-2   | Rel.reprod. s.d.:   | 13,30%                                |
| Rel.target s.d.: | 10,00%       | Reference value:    | 80,1 µg/m³                            |
|                  |              | Range of tolerance: | 69,1 - 103,7 µg/m³ ( Z-Score  ≤ 2,00) |




## Summary results

|                         |                     |                            |   |
|-------------------------|---------------------|----------------------------|---|
| <b>Measurand:</b>       | <b>Benzaldehyde</b> | <b>Mean:</b>               | <b>26,8 µg/m³</b>                               |
| <b>Sample:</b>          | <b>2</b>            | <b>Reprod. s.d.:</b>       | <b>4,4 µg/m³</b>                                |
| <b>Method:</b>          | <b>ISO 5725-2</b>   | <b>Rel.reprod. s.d.:</b>   | <b>16,34%</b>                                   |
| <b>Rel.target s.d.:</b> | <b>10,00%</b>       | <b>Reference value:</b>    | <b>24,3 µg/m³</b>                               |
|                         |                     | <b>Range of tolerance:</b> | <b>21,4 - 32,1 µg/m³ ( Z-Score  &lt;= 2,00)</b> |




## Summary results

|                  |            |                     |                                      |
|------------------|------------|---------------------|--------------------------------------|
| Measurand:       | Benzene    | Mean:               | 83,0 µg/m³                           |
| Sample:          | 2          | Reprod. s.d.:       | 11,4 µg/m³                           |
| Method:          | ISO 5725-2 | Rel.reprod. s.d.:   | 13,78%                               |
| Rel.target s.d.: | 10,00%     | Reference value:    | 82,7 µg/m³                           |
|                  |            | Range of tolerance: | 66,4 - 99,6 µg/m³ ( Z-Score  ≤ 2,00) |




## Summary results

|                         |                   |                            |   |
|-------------------------|-------------------|----------------------------|---|
| <b>Measurand:</b>       | <b>Cumene</b>     | <b>Mean:</b>               | <b>51,5 µg/m³</b>                           |
| <b>Sample:</b>          | <b>2</b>          | <b>Reprod. s.d.:</b>       | <b>5,1 µg/m³</b>                            |
| <b>Method:</b>          | <b>ISO 5725-2</b> | <b>Rel.reprod. s.d.:</b>   | <b>9,97%</b>                                |
| <b>Rel.target s.d.:</b> | <b>10,00%</b>     | <b>Reference value:</b>    | <b>47,6 µg/m³</b>                           |
|                         |                   | <b>Range of tolerance:</b> | <b>41,2 - 61,8 µg/m³ ( Z-Score  ≤ 2,00)</b> |




## Summary results

|                         |                   |                            |   |
|-------------------------|-------------------|----------------------------|---|
| <b>Measurand:</b>       | <b>n-Decane</b>   | <b>Mean:</b>               | <b>77,5 µg/m³</b>                               |
| <b>Sample:</b>          | <b>2</b>          | <b>Reprod. s.d.:</b>       | <b>10,9 µg/m³</b>                               |
| <b>Method:</b>          | <b>ISO 5725-2</b> | <b>Rel.reprod. s.d.:</b>   | <b>14,02%</b>                                   |
| <b>Rel.target s.d.:</b> | <b>10,00%</b>     | <b>Reference value:</b>    | <b>79,4 µg/m³</b>                               |
|                         |                   | <b>Range of tolerance:</b> | <b>62,0 - 93,0 µg/m³ ( Z-Score  &lt;= 2,00)</b> |




## Summary results

|                  |            |                     |  |
|------------------|------------|---------------------|--|
| Measurand:       | n-Heptane  | Mean:               | 73,2 µg/m <sup>3</sup>                           |
| Sample:          | 2          | Reprod. s.d.:       | 9,6 µg/m <sup>3</sup>                            |
| Method:          | ISO 5725-2 | Rel.reprod. s.d.:   | 13,09%   |
| Rel.target s.d.: | 10,00%     | Reference value:    | 70,6 µg/m <sup>3</sup>                           |
|                  |            | Range of tolerance: | 58,5 - 87,8 µg/m <sup>3</sup> ( Z-Score  ≤ 2,00) |




## Summary results

|                         |                   |                            |  |
|-------------------------|-------------------|----------------------------|--|
| <b>Measurand:</b>       | <b>p-Xylene</b>   | <b>Mean:</b>               | <b>59,0 µg/m<sup>3</sup></b>                           |
| <b>Sample:</b>          | <b>2</b>          | <b>Reprod. s.d.:</b>       | <b>6,9 µg/m<sup>3</sup></b>                            |
| <b>Method:</b>          | <b>ISO 5725-2</b> | <b>Rel.reprod. s.d.:</b>   | <b>11,70%</b>  |
| <b>Rel.target s.d.:</b> | <b>10,00%</b>     | <b>Reference value:</b>    | <b>59,2 µg/m<sup>3</sup></b>                           |
|                         |                   | <b>Range of tolerance:</b> | <b>47,2 - 70,8 µg/m<sup>3</sup> ( Z-Score  ≤ 2,00)</b> |



## Summary results


|                         |                   |                            |   |
|-------------------------|-------------------|----------------------------|---|
| <b>Measurand:</b>       | <b>Toluene</b>    | <b>Mean:</b>               | <b>48,8 µg/m³</b>                           |
| <b>Sample:</b>          | <b>2</b>          | <b>Reprod. s.d.:</b>       | <b>7,8 µg/m³</b>                            |
| <b>Method:</b>          | <b>ISO 5725-2</b> | <b>Rel.reprod. s.d.:</b>   | <b>15,97%</b>                               |
| <b>Rel.target s.d.:</b> | <b>10,00%</b>     | <b>Reference value:</b>    | <b>47,6 µg/m³</b>                           |
|                         |                   | <b>Range of tolerance:</b> | <b>39,1 - 58,6 µg/m³ ( Z-Score  ≤ 2,00)</b> |





# Sample chart of Z-scores

Sample 2



## Summary of laboratory test results

| Laboratory                                 | 1,2,3-Trimethylbenzene | 4-Methyl-2-pentanone | alpha-Pinene      | Benzaldehyde      | Benzene           | Cumene            | n-Decane          |
|--|------------------------|----------------------|-------------------|-------------------|-------------------|-------------------|-------------------|
| Unit                                       | µg/m <sup>3</sup>      | µg/m <sup>3</sup>    | µg/m <sup>3</sup> | µg/m <sup>3</sup> | µg/m <sup>3</sup> | µg/m <sup>3</sup> | µg/m <sup>3</sup> |
| 1  | < 0.50                 | < 0.50               | < 0.50            | < 2.00            | < 0.50            | < 0.50            | < 0.50            |
| 24   | 1.12                   | 0.00                 | 0.67              | 5.90              | 0.00              | 0.64              | 0.32              |
| 26   | 0.20                   | 0.20                 | 0.15              | 0.50              | 0.60              | 0.07              | 0.30              |
| 39   | < 0.50                 | < 0.50               | 1.85              | 4.05              | < 0.50            | < 0.50            | 1.90              |
| 44   | 0.17                   | 0.00                 | 0.15              | 2.17              | 1.90              | 0.05              | 0.00              |
| 66   | < 1.00                 | < 2.00               | < 1.00            | 2.20              | < 0.50            | < 1.00            | < 1.00            |
| 95   | 0.60                   | 3.05                 | 0.07              | 2.03              | 0.00              | 0.48              | 0.50              |
| 147  | 0.00                   | 0.00                 | 0.00              | 0.00              | 0.00              | 0.00              | 0.00              |
| 152  | 0.00                   | 0.00                 | 0.00              | 0.00              | 0.00              | 0.00              | 0.00              |
| 180  | 0.00                   | 0.00                 | 0.00              | 0.00              | 0.00              | 0.00              | 0.00              |
| 209  | < 1.00                 | < 1.00               | < 1.00            | < 1.00            | < 1.00            | < 1.00            | < 1.00            |
| 213  |                        |                      |                   | 2.52              |                   |                   |                   |
| 228  | < 1.00                 |                      |                   |                   | 2.50              | < 1.00            | < 1.00            |
| 230  | < 1.00                 | < 2.00               | < 2.00            | < 2.00            | < 1.00            | < 1.00            | < 2.00            |
| 235  | < 0.10                 | < 0.10               | < 0.10            | < 0.10            | < 0.10            | < 0.10            | < 0.10            |
| 242  | 0.00                   | 0.00                 | 0.00              | 0.00              | 0.00              | 0.00              | 0.00              |
| 265  | 0.00                   | 0.00                 | 0.00              | 0.00              | 0.00              | 0.00              | 0.00              |
| 284  | 0.00                   | 0.00                 | 0.00              | 0.00              | 0.00              | 0.00              | 0.00              |
| 289  | 0.09                   | < 2.00               | 0.02              | 0.31              | 0.12              | 0.04              | 0.12              |
| 517  | < 10.00                | < 10.00              | < 10.00           | < 10.00           | < 10.00           | < 10.00           | < 10.00           |
| --   | --                     | --                   | --                | --                | --                | --                | --                |
| No. of laboratories that submitted results | 19                     | 18                   | 18                | 19                | 19                | 19                | 19                |

| Laboratory                                 | n-Heptane         | p-Xylene          | Toluene           |
|--|-------------------|-------------------|-------------------|
| Unit                                       | µg/m <sup>3</sup> | µg/m <sup>3</sup> | µg/m <sup>3</sup> |
| 1  | < 0.50            | < 0.50            | < 0.50            |
| 24   | 0.00              | 0.52              | 0.28              |
| 26   | 0.40              | 1.10              | 2.95              |
| 39   | < 0.50            | 1.30              | 1.30              |
| 44   | 0.22              | 0.13              | 0.75              |
| 66   | < 1.00            | 5.70              | < 0.50            |
| 95   | 0.00              | 0.59              | 0.78              |
| 147  | 0.00              | 0.00              | 0.00              |
| 152  | 0.00              | 0.00              | 0.00              |
| 180  | 0.00              | 0.00              | 0.00              |
| 209  | < 1.00            | < 1.00            | < 1.00            |
| 228  | < 1.00            | 1.00              | 1.50              |
| 230  | < 2.00            | < 1.00            | 3.00              |
| 235  | < 0.10            | < 0.10            | < 0.10            |
| 242  | 0.00              | 0.00              | 0.00              |
| 265  | 0.00              | 0.00              | 0.00              |
| 284  | 0.00              | 0.00              | 0.00              |
| 289  | < 2.00            | 0.07              | 0.03              |
| 517  | < 10.00           | < 10.00           | < 10.00           |
| -  | --                | --                | --                |
| No. of laboratories that submitted results | 19                | 19                | 19                |

## Questions and Answers

| Participant | Sample carrier   |
|-------------|--|
| 1           | Tenax TA   |
| 23          | Tenax Rohre von Markes   |
| 24          | Gerstel Tenax TA   |
| 26          | Tenax TA (Markes)  |
| 39          | Tenax-Röhrchen ( Vom Labor)  |
| 44          | Tenax Markes   |
| 66          | Tenaxröhrchen  |
| 95          | Tenax GR   |
| 126         | Tenax TA   |
| 147         | Carbotrap 349  |
| 152         | Tenax TA v. Fa. Supelco  |
| 180         | TENAX  |
| 206         | Tenax TA CAMSCO  |
| 209         | Thermodesorptionsröhrchen auf Mehrbettssystem (Tenax TA und Carbopack X)   |
| 213         | TDU Röhrchen mit TENAX Ta, Hersteller:- Markes   |
| 228         | Tenax  |
| 230         | Tenax TA, Fa. Gerstel  |
| 233         | Edelstahlröhrchen mit Tenax TA von Markes  |
| 235         | stainless steel TD tube, O.D. × L 1/4 in. × 3 1/2 in., 60-80 mesh, Tenax TA, Supelco ( <a href="https://www.sigmaaldrich.com/DE/de/product/supelco/30133u">https://www.sigmaaldrich.com/DE/de/product/supelco/30133u</a> ) |
| 242         | Tenax TA, Fa. Gerstel  |
| 265         | Tenax TA   |
| 284         | Tenax TA v. Fa. Supelco  |
| 289         | Tenax TA   |
| 517         | Tenax TA (Fa. Supelco)   |

| Participant | Sampling pump                     | Flow rate     | Flow rate measurement   | Sampling time |
|-------------|-----------------------------------|---------------|-------------------------|---------------|
| 1           | SKC PCXR8, AirChek2000, GSA 350ex | ca. 0,1 l/min | Massflow meter TSI 4100 | 10 und 30 min |
| 23          | Bivoc                             | 100 ml/min    |                         | 50 min        |
| 24          | BIVOC2V2                          | 0,1-0,2 L/min | BIVOC2V2                | 10-40min      |

**Proficiency testing scheme VOC with own sampling 2023**

| Participant | Sampling pump                          | Flow rate                           | Flow rate measurement            | Sampling time               |
|-------------|--|-------------------------------------|----------------------------------|-----------------------------|
| 26          | Gilian LFS-113 DC                      | 0,067 - 0,070 l/min                 | Gilibrator 3                     | 30                          |
| 39          | Gilian Air Plus                        | 0,1L/min                            | Gilian Gilibrator 3              | 30                          |
| 44          | BIVOC2, Fa. Holbach                    | 0,15 L/min.                         | In der Pumpe integriert          | 20                          |
| 66          | SKC AirChek TOUCH                      | 50 ml/min                           | Analyt-MTC                       | 60                          |
| 95          | GSA SG350ex                            | 0,1 L/min                           | DryCal DC-Lite                   | 30 min, 45 min, 60 min      |
| 126         | Aircon 2                               | M1: 200 mL/min; M2: 100 mL/min      | DryCal Defender 530              | 30                          |
| 147         | GilAir plus                            | 250 ml/min                          | Bios Defender                    | 80 min                      |
| 152         | Turzer                                 | 0.2 l/min                           | interner Massflow controller     | 15                          |
| 180         | KNF Laboport Mini-Laborpumpen          | 0.15 l/min                          | red-y compact series by vögtlin  | 81-90 min                   |
| 206         | SKC 224-PCEX8                          | 100 ml/Min.                         | Defender 510 BIOS                | 30 Minuten                  |
| 209         | GilAir Plus Personal Air Sampling Pump | 190 mL/Min                          | TSI 4100                         | 15                          |
| 213         | BIVOC2                                 | 0,2ml/min                           | BIVOC2                           | 25min                       |
| 228         | Hersteller Holbach Model BiVOC2        | 0,2 l/min                           | Hersteller Holbach Model BiVOC2  | Ca. 25 min                  |
| 230         | BiVOC, Fa. Holbach                     | 0,15 ml/ml                          | TSI 4140 F                       | 20                          |
| 233         | DESAGA GS301 und Gilian GilAir Plus    | 0,1 L/min                           | Aalborg GFM17                    | 20                          |
| 235         | GSA Messgerätebau, GSA 350ex           | 167 ml/min (BW, M1); 100ml/min (M2) | GSA Messgerätebau, GSA 350ex     | BW 45; M1 60 Min; M2 30 Min |
| 242         | LFS-113 DC                             | 100ml/min                           | Defender 510                     | 20                          |
| 265         | BiVOC V2                               | 0,1 L/min                           | Massenflussmesser von Bronkhorst | 20                          |
| 284         | GilAir                                 | 200 ml/min                          | Gilibrator                       | 15 Minuten                  |
| 289         | Gilair Plus                            | 6 L                                 | Gilibrator                       | 60                          |
| 517         | SKC Universal                          | 50 ml/min                           | Strömungsmesser Analyt           | 60                          |

| Participant | Analytical method                             | Gas chromatograph (GC)                             |
|-------------|---|--|
| 1           | DIN ISO 16000-6 (03-2022)                     | Shimadzu QP-2020                                   |
| 23          | 16000-6                                       | GC Shimadzu 2010 Plus                              |
| 24          | DIN ISO 16000-6                               | Agilent GC 7890 B                                  |
| 26          | DIN ISO 16000-6, DIN EN 16516                 | Agilent GC 7890B                                   |
| 39          | DIN ISO 16000-6 (03-2022)                     | Shimadzu QP-2020                                   |
| 44          | DIN ISO16000-6, DIN EN ISO 16017-1, EPA to 17 | HPGC MS D-13 GC-System Argelent 8890+ 5977B GC MSD |
| 95          | DIN EN ISO 16017-1                            | Agilent 8890                                       |
| 126         | DIN EN ISO 16000-6                            | ThermoFisher Trace GC Ultra                        |
| 147         | ISO 16000-6                                   |  |

Proficiency testing scheme VOC with own sampling 2023

| Participant | Analytical method             | Gas chromatograph (GC)   |
|-------------|-------------------------------|--|
| 152         | DIN ISO 16000-6               | Agilent 7890B  |
| 180         | DIN 16000-6:2022-03           | Shimadzu GC-MS QP 2010 Ultra bzw . Shimadzu GC-MS-QP2020, 60 m Säule |
| 206         | DIN ISO 16000-6               | AGILENT 7890 A   |
| 209         | DIN EN ISO 16000-6            | -  |
| 213         | ISO 16000-6                   | Agilent 8890 GC System   |
| 228         | DIN ISO 16000-6               | Agilent  |
| 230         | DIN ISO 16000-6               | Thermo Scientific GC Trace 1610 + ISQ 7610                           |
| 233         | DIN ISO 16000-6               | Agilent GC 7890A   |
| 235         | ISO 16017-1                   | PerkinElmer: GC: Clarus 600; MS: Clarus SQ8                          |
| 242         | nein                          | Agilent GC 8890  |
| 265         | DIN ISO 16000-6               | Shimadzu QP2020  |
| 284         | DIN ISO 16000-6               | Agilent 7890B  |
| 289         | DIN ISO 16000-6, DIN EN 16516 | Agilent GC 7890B   |
| 517         | DIN ISO 16000-6               | Agilent 7890B  |

| Participant | Thermal desorber              | Desorption temperature | Cryo trap                         | Carrier gas |
|-------------|-------------------------------|------------------------|-----------------------------------|-------------|
| 1           | Shimadzu TD-30R               | 220°C                  | - 30 °C auf 220°C                 | Helium      |
| 23          | Shimadzu TD30-R               | 250 C°                 | -20 bid 250 C°                    | Helium      |
| 24          | Gerstel TDS 3                 | 260°C                  | -150°C / 280°C                    | Helium      |
| 26          | Markes TD100                  | 280° C                 | -20 / 315° C                      | Helium      |
| 39          | Shimadzu TD-30R               | 220°C                  | -30°C auf 220°C                   | Helium      |
| 44          | Turbo Matrix 650 Perkin Elmer | 280°C                  | -20°C, 40°C/sek.                  | Helium      |
| 95          | Gerstel TD3.5+                | 270°C                  | -145°C - 12°C/sec - 300°C - 3 min | Helium      |
| 126         | Markes TD100XR                | 250°C                  | -20°C; Heiztemperatur: 250°C      | Helium      |
| 152         | Gerstel TD 3.5.+              | 265°C                  | -30°C / 270°C                     | Helium      |
| 180         | Markes TD100-xr               |                        |                                   |             |
| 206         | PERKIN-ELMER TurboMatrix 650  | 275°C                  | -30°C bis 280°C                   | Helium      |
| 209         | Shimadzu TD-20                | 240                    | -18 °C / +260 °C                  | He          |
| 213         | Markes TD100-xr               | 260°C                  | -30°C, 300°C                      | Helium      |
| 228         | Markes Unit 2                 | 280°C                  | 5°C / 60°C / min auf 280°C        | Helium      |
| 230         | TDS 3, Fa. Gerstel            | 40°C-260°C             | -30°C bis 260°C                   | Helium      |
| 233         | Markes TD100-xr               | 280°C                  | -10°C auf 320°                    | Helium      |

**Proficiency testing scheme VOC with own sampling 2023**

| Participant | Thermal desorber     | Desorption temperature | Cryo trap       | Carrier gas                                 |
|-------------|----------------------|------------------------|-----------------|---|
| 235         | ATD: Turbomatrix 650 | 280°C                  | -30°C           | Helium 5.0, zusätzlich Aktivkohle-filtriert |
| 242         | TDS 3                | 280°C                  | -150°C; 60°C/s  | Helium                                      |
| 265         | Shimadzu TD30        | 290 °C                 | -17 °C / 290 °C | Helium                                      |
| 284         | Gerstel TD 3.5.+     | 265°C                  | -30°C / 270°C   | Helium                                      |
| 289         | Markes TD100         | 280°C                  | -20 / 315       | Helium                                      |
| 517         | Gerstel TD 3.5.+     | 265 °C                 | -30 °C / 270 °C | Helium                                      |

| Participant | Carrier gas flow                                    | Analytical column  | Detector                          |
|-------------|---|--|-----------------------------------|
| 1           | 1,2 ml/min  | Restek RTX-5ms   | MS und FID                        |
| 23          | Column Flow : 2,42 ml/min, Total Flow : 33,1 ml/min | Rxi 5 Sil ms   | MS Shimadzu QP-2020 /Quadropol-MS |
| 24          | 1,3ml/min   | Agilent Ultra 2  | Agilent MSD 5977B                 |
| 26          | 0,5   | Restek Rxi-5Sil MS, 20m x 0,18mm id x 0,36µm df                | MS                                |
| 39          | 1,2ml/min   | Restek RTX-5ms   | MS und FID                        |
| 44          | 3,1 ml/min über Thermodesorber                      | RTX 502.2 60m x 250 µm x 1,4 µm                                | Massenspektrometer                |
| 95          | 0,8 mL/min  | Rtx-502.2; 40m x 0,18mm ID; 1µmFD                              | Agilent 5975C MSD                 |
| 126         | 1 mL/min  | Rxi-624 Sil MS 60 m x 0,25 mm, 1,4 µm Filmdicke                | MSD ISQ 7000                      |
| 152         | 1,2ml/min   |  | MSD 5977 B, Agilent               |
| 206         | 1,5 ml/Min.   | RESTEK Rtx-5Sil MS   | LECO Pegasus 4 D                  |
| 209         | 45 ml/min   | DB-5MS 60 m - 0,25 µm - 0,25 mm (Agilent)                      | MS QP2010 Ultra, Shimadzu         |
| 213         | 1.3ml/min   | ULTRA HP   | Agilent 5977B GC/MSD              |
| 228         | 1.3 ml/min  | HP Ultra 2   | Agilent MS                        |
| 230         | 06 ml/min   | Restek Rxi 1 MS, 60 m  | MS                                |
| 233         | 1,5 mL/min  | HP-5ms 5% Phenyl (30 m; ID: 0,25 mm, Filmdicke 0,25 µm)        | Agilent 5975C MSD                 |
| 235         | 1 ml/min  | PerkinElmer Elite 5-MS, l = 60 m, ID = 0,25 mm, film = 0,25 µm | MS: Clarus SQ8                    |
| 242         | 1,3ml/min   | HP-5 Ultra 2   | Massenspektrometer 5977C          |
| 265         | 2,46  | Agilent VF-5MS   | MS                                |
| 284         | 1,2ml/min   | MSD 5977 B, Agilent  | MSD 5977 B, Agilent               |
| 289         | 0.5   |  |                                   |
| 517         | 1,2   |  | MSD 5977 B, Agilent               |

## Proficiency testing scheme VOC with own sampling 2023

| Participant | Data evaluation   |
|-------------|---|
| 1           | Identifizierung: Retentionszeit + Massenspektrum; Quantifizierung: MS interne Standardmethode, FID externe Standardmethode  |
| 23          | Quantifizierung über Vergleichstandards (substanzspezifisch)  |
| 24          | 2-Punktkalibrierung externer Standards, Identifikation via Spektrum   |
| 26          | Identifizierung und Quantifizierung mittels GC-MSD und Toluol d8 als interner Standard, 12 Punkt Kalibrierung aller Analyten (1-500 ng abs.)                                      |
| 39          | Identifizierung: Retentionszeit + Massenspektrum; Quantifizierung: MS interne Standardmethode, FID externe Standardmethode  |
| 44          | externe Kalibrierung (Fläche)   |
| 95          | Quantifizierung über internen Standard, Identifizierung über MS-Spektrum und RT   |
| 126         | Quantifizierung über TIC; Ausnahme alpha-Pinen: quan via m/z 93 und ident via m/z 91  |
| 152         | Identifizierung: stoffspezifisch, Vergleichsspektrum+Retentionszeit   |
| 206         | externer Standard   |
| 209         | Quantifizierung: Substanzspezifisch mittels ext. Kalibrierreihe (5-Punkt- Kalibrierung), Identifizierung: Referenzmassenspektrum und Retentionszeit                               |
| 213         | Quantifizierung der spezifischen Substanzen mit Drei-Punkt-Kalibrierung, Identifizierung mit Retentionszeit und Ionenfragmenten   |
| 228         | Einzelauswertung mit spezifischen Massen  |
| 230         | Referenzstands, eigene Belegung, Retentionszeit, MS   |
| 233         | externe 6-Punkt-Kalibrierung mit ISTD (Cyclooctan)  |
| 235         | Quantifizierung über Toluene-D8 internen Standard (100ng); Identifikation über MS-NIST Datenbank und Standardsubstanzen (Supelco P/N 49148-U, P/N 28220, Sigma-Aldrich P/N B1334) |
| 242         | Quantifizierung über externe Mehrpunktkalibration der Einzelsubstanzen, Identifikation über Massenspektrometrie   |
| 265         | über stoffspezifische Kalibration   |
| 284         | stoffspezifisch, Vergleichsspektrum + Retentionszeit  |
| 289         | GC-MSD and Toluol d8 as internal standard, 12 point calibration of all analytes (1-500ng abs.)  |
| 517         | stoffspezifisch, Vergleichsspektrum und Retentionszeit  |

| Participant | Recovery rate | Date of analysis          |
|-------------|---------------|---------------------------|
| 1           | nein          | 22.09.2023                |
| 23          | nein          | 12.09.2023+ 22.09.2023    |
| 24          | nein          | 16.-17.10.2023            |
| 26          | nein          | 13.09.2023                |
| 39          | nein          | 22.09.2023                |
| 44          | nein          | 15.09.2023                |
| 66          |               | 21.09.2023                |
| 95          | nein          | 21.09.2023                |
| 126         | Nein          | 19.09.2023 bis 13.10.2023 |



### Proficiency testing scheme VOC with own sampling 2023

---

| Participant | Recovery rate | Date of analysis               |
|-------------|---------------|--------------------------------|
| 152         | nein          | 21.9.23                        |
| 180         |               | 25.09.-16.10.2023              |
| 206         | ja            | 02.10.2023                     |
| 209         | Nein          | 18.09. - 04-10.2023            |
| 213         | nein          | 10.10.2023 & 12.10.2023        |
| 228         | Nein          | 18.09.2023                     |
| 230         | nein          | ab dem 15.09.2023              |
| 233         | nein          | 13.09.-16.10.2023              |
| 235         | nein          | 2 bzw . 6 Tage nach Probenahme |
| 242         | Nein          | 14.09.2023                     |
| 265         | nein          | 18.09.2023                     |
| 284         | nein          | 21.9.23                        |
| 289         | no            | 19.09.2023                     |
| 517         | nein          | 21.09.2023                     |